



DIVISION OF INDUSTRY SERVICES
Plumbing Product Review
P.O. Box 2658
Madison, Wisconsin 53701-2658
TTY: Contact Through Relay

Scott Walker, Governor
Dave Ross, Secretary

August 22, 2013

MEDIVATORS
RENAL SYSTEMS
REGULATORY AFFAIRS
14605 28TH AVENUE NORTH
MINNEAPOLIS MN 55441

Re: Description: HEALTH CARE PLUMBING APPLIANCE
Manufacturer: MEDIVATORS
Product Name: RENATRON AND RENATRON II
Model Number(s): RS8300 AND RS8330
Product File No: 20130147

The specifications and/or plans for this plumbing product have been reviewed and determined to be in compliance with chapters SPS 382 through 384, Wisconsin Administrative Code, and Chapters 145 and 160, Wisconsin Statutes.

The Department hereby issues an approval based on the Wisconsin Statutes and the Wisconsin Administrative Code. This approval is valid until the end of August 2018.

This approval supersedes the approval issued on October 22, 2008 under product file number 20080591.

This approval is contingent upon compliance with the following stipulation(s):

- A reduced pressure zone (ASSE 1013) or an air gap (ASME A112.1.2) shall be installed on the influent water line serving these devices.
- An air gap (ASME A112.1.2) shall be installed on the effluent water line serving these devices.

The department is in no way endorsing this product or any advertising, and is not responsible for any situation which may result from its use.

Sincerely,

Glen W. Schlueter
Environmental Engineer - Plumbing Product Reviewer
Department of Safety and Professional Services
Division of Industry Services
Bureau of Technical Services
(608) 267-1401 Phone
(608) 267-9723 Fax
glen.schlueter@wi.gov E-mail